SHAMPOO BASIN

Réf.N° : 16 350 40

Instructions

We thank you for choosing the shampoo basin distributed by COMED and we hope that you will be satisfied. In any case you can join us for more information at our adress below or contact us by email at info@comed.fr

1. Unfold the shampoo basin to make it seamless. Inflate it thanks to a pump (not included, COMED item no. 16 350 42) or the mouth.
2. Place the basin under the patient head while respecting the position for the nape.
3. Place the end of the drain tube in a container placed lower than the shampoo basin.
4. After the shampoo, rinse with water (portable water bag with shower head in option, COMED item no. 16 350 45)
5. To drain off the water from the basin, open the drain in the bottom of the basin. You can afterards remove the basin from patient's head.
6. Make sure of well rinsing and drying the basin after use. For a better drying, keep it inflated for some hours.

Sté. COMED , Parc des forges , 8 Rue Louise Michel 67200 STRASBOURG -FRANCE ,
Tél. : 03 88 77 13 00 , Fax : 03 88 77 00 70 Site : www.comed.fr
E-mail : info@comed.fr

QUALITE\Fiche Technique TCE1 TCE2 TCE3\TCE3\Dossier technique et sécurité TCE3\Bac à shampooing\Mode d'emploi 16 350 40
LH 24.04.13